


Visual Melting Range Apparatus

(MR-VIS⁺)



LABINDIA
ANALYTICAL
WHERE COMMITMENT IS CULTURE





VISUAL MELTING RANGE APPARATUS

Visual Melting Range Apparatus is the latest, state-of-the-art microcontroller based instrument for determining Melting Range of white as well as coloured samples in powdered form. Around 5mg of sample is uniformly spread on a glass slide and covered with a coverslip. The slide is then placed on a specially designed uniformly heated round furnace and subjected to the heating rate specified by the user.

A magnified image of the sample can be observed on a video monitor screen via a CCD camera. The effect of heat on the sample can be clearly seen. The change in physical appearance of the sample with reference to temperature is recorded and the start of melting & end of melting are determined automatically. This automatic detection of start and end of melting can be easily verified by observing the melting process on the video screen.

A representation of the entire melting process in the form of a graphics printout is available for authentication as illustrated. This instrument also has a manual operation facility to lock the start of melting and end of melting. In Heat and Cool mode the Start Temperature of Melting as well as Solidification Temperature of sample can be locked manually.

The apparatus can also be used to detect melting point by capillary method by simply attaching capillary conversion accessory (optional).

Applications

- Organic / Inorganic Chemicals
- Pharmaceuticals
- Petrochemicals
- Dyes
- Plastics / Polymers
- Fibers / Textiles etc.

PRINTOUT

Various printouts are available as follows :

CUSTOMER NAME :

Calibration REPORT

Instrument : LABINDIA MR-Vis. 15:57:21 11/11/2012
Inst. Sr No. : 2
Validation Date : 11/11/2012
Validation Time : 14:55

Data of Standards used:

Ref. Std. 1

Std Name : Vanillin
Ref Id.No. : A
Melting Range°C : 81.0 - 83.0

Ref. Std. 2

Std Name : Caffeine
Ref Id.No. : A
Melting range°C : 235.5-237

Validation Analysis

Sample Preparation : Capillary
Operation : Auto
Set Temp : 81.0°C
No.of Readings : 3
M.P. (1) : 82.0°C
M.P. (2) : 81.6°C
M.P. (3) : 81.8°C

Set Temp : 235.5°C
No. of Readings : 3
M.P (1) : 236.0°C
M.P (2) : 236.6°C
M.P. (3) : 236.3°C

Avg M.P : 81.8°C

Avg M.P. : 236.3°C

Name : _____

Signature : _____

Date : _____

15:58:16 11/11/2012


CUSTOMER NAME :

Melting Range Analysis Report

Inst. Sr.No:2
Run No. :8 Run Date & Time:11/11/2012 15:54
ID. No. :D Sample No. :D
Prog No. :2 Sample Prep. :Slide
Sample Name:VANILLIN Mode :Heat
Set Temp. :81.0°C Heat Rate :1.0°C/MIN
Operation :Automatic
Results :
Start Temp :81.3°C End Temp :82.4°C
M.P. :81.5°C

Report printed - 15:57:53 11/11/2012
Analysis done on LABINDIA MR-Vis.

Run No. : 8
Sample Name : Vanillin
Sample No. : D



Temp°C X Axis Scale: 23.0 L.units/div
Y Axis Scale: 0.660°C/div

Start Temp : 81.3°C End temp : 82.4°C

Features

- Advanced microcontroller based user-friendly, state-of-the-art design with alphanumeric splash waterproof polyester soft keys for keyboard. User interactive software in dialogue mode for ease of operation with protection against invalid entries.
- Visual Melting Range Apparatus is used for detecting the melting range / melting point of substances.
- Alphanumeric entries of sample name, sample number, identification number with date and time for authentication. Daily Auto Incremented Run number and factory entered CUSTOMER NAME and Instrument Sr. No. on report printouts makes system foolproof and GLP compliant.
- Non-volatile memory storage of 20 methods with parameters. Last run result can be viewed or printed.
- Built-in 2-point Automatic calibration.
- Calibration data with date & time stamp for authentication.
- User selectable operating modes :
 - a) In AUTO detection mode, start of melting and end of melting are automatically detected by photo sensing the melting process via CCD camera.
 - b) In MANUAL detection mode, start of melting and end of melting can be locked with a key-press by user.
- Built-in Real time clock (RTC) for time display and on report printout.
- Intelligent Lamp Intensity Control with Soft Start.
- Selectable Report Format, complying with GLP requirements:
 - a) Report giving parameter and result.
 - b) Report of calibration data.
 - c) Graphics report of temperature v/s L.units with markings of start of melting and end of melting.
 - d) Report of program parameters for 20 programs.
- Error indication helps the user to trace the problem.
- Melting RUN can be started with last run parameters.
- Facility to abort RUN by single key.
- Escape and Back key for better user interaction.
- STOP key for achieving isothermal condition during manual RUN.
- Printer compatibility : a) Dotmatrix (standard) b) DeskJet (optional)
- **Optional**
 - i) Capillary conversion attachment.
 - ii) Compliance with USP melting range procedure for class Ia.
 - iii) IQ, OQ, PQ, documents available.
- **MR-VIS⁺**
 - i) Software for Image Capture on PC.
 - ii) OSD Adapter for authentication of PC captured image.

Specifications

Control	Microcontroller-based (Advanced version of microprocessor).
Furnace construction	Round Block.
Temperature sensor	PT100 (RTD)
Temperature Range	Ambient + 5°C to 350 °C
Temperature Readability	0.1°C
Heating Rates	0.2°C/min, ~ 12.0°C/min,
Max. Cooling Time from 350°C to ambient (25 °C)	25 minutes (approx.)
Accuracy of Detection of melting temperature	
a) Ambient + 5°C to 200°C - ± 0.5°C	
b) 200°C to 300°C - ± 0.8°C	
c) Above 300°C - ± 1.4°C	
Sampling	
a) Sample Size - 5 mg (Approx.)	
b) Sample Holder i) Glass slide < = 1mm ± 0.02 mm thick ii) Glass capillary tube with one end sealed (optional accessory)	
Visual Image	10X magnified image displayed on the monitor via CCD camera.
Keyboard	Alphanumeric splash water proof polyester soft keys.
Display	
a) 20 x 2 line back lighted liquid crystal display (LCD).	
b) Portable B&W monitor to view magnified image of sample.	
c) Colour CCD with colour monitor (optional).	
Data storage	Non-volatile memory
Program storage	20 methods with parameters
Printer output	Centronix parallel port
Report format	GLP compliant - Selection of formats including result report, graphics, calibration report and method parameters.
Power	230 VAC ±10%, 50Hz.
Modes of detection	AUTO & MANUAL
Environmental Operating Conditions	
a) Operation Indoor	
b) Temperature Ambient to 45°C	
c) Humidity 5 to 90% non-condensing.	

LABINDIA

ANALYTICAL
WHERE COMMITMENT IS CULTURE

www.labindia-analytical.com / For Enquiries: sales.mfd@labindia.com

Labindia Analytical Instruments Pvt. Ltd.

FACTORY: Plot No. EL-72, Electronic Zone, TTC Industrial Area, Thane Belapur Road, Navi Mumbai - 400 705. Tel.: +91-22-2762 6660/61/62/63/65163480/81
Fax: +91-22-27626664 E-mail: lipfactory@vsnl.net

APPLICATION & TRAINING CENTRE: R-909, TTC Industrial Area, Thane Belapur Road, Rabale, Navi Mumbai - 400 701. Tel.: +91-22-2760 6955 Fax: +91-22-2760 6706

HEAD OFFICE:

Thane: 201, Nand Chambers, LBS Marg, Near Vandana Cinema, Thane (W) - 400 602. Tel.: +91-22-2598 6250 / 2598 6282 Fax: +91-22-2541 0420 / 2533 5940
E-mail: sales.mfd@labindia.com

REGIONAL OFFICES:

Delhi: +91-11-4330 6001 / 10

Chennai: Tel.: +91-44-2434 7008 / 2432 0352

Kolkatta: Tel.: +91-33-2466 1396 / 2466 3362

BRANCH OFFICES:

Bangalore: Tel.: +91-80-2323 0923 / 21

Lucknow: Tel.: +91-522-2346 535 / 2346 496

Hyderabad: Tel.: +91-40-6550 4133 / 2331 3300

Gurgaon: Tel.: +91-124-2843 300 / 2843 600

Tiruvanthapuram: Tel.: +91-471-2324 064 / 2320 082

Vadodara: Tel.: +91-265-2342 929

Pune: Tel.: 098220 43622

Chandigarh: Tel.: +91-172-4090 001-07 / 4090 009



Striving to become the best individuals, we endeavour to foster the best team. Performing sensibly, we try to achieve the best efficiency. Working innovatively, we seek to make the best products. Listening patiently, we excel to offer the best service. So, no matter what your needs are, **come to us, GET THE BEST**

LABINDIA reserves the right to change specifications without notice as part of its continuous programme of product development.